

Mouse embryo test report

Test products:

Product name: Petri dish for assisted reproduction

Representative model: MC2010

Product batch number: 2330200003

Production date: July 26, 2023

Test standard:

YY/T 1434-2016 "In vitro mouse embryo test of medical apparatus for human in vitro assisted reproductive technology"

Test time:

Sep 7, 2023 - Sep 11, 2023

Test results:

Blastocyst formation rate $\geq 80\%$

Wuhan Huchuang United Technology Co., Ltd.

Quality Department

Sep 11, 2023