



Mouse embryo test report

Test products:

Product name: Petri dish for assisted reproduction

Representative model: MC2011

Product batch number: 2330200004

Production date: July 27, 2023

Test standard:

YY/T 1434-2016 "In vitro mouse embryo test of medical apparatus for human in vitro assisted reproductive technology"

Test time:

Sep 2, 2023 - Sep 6, 2023

Test results:

Blastocyst formation rate $\geq 80\%$

Wuhan Huchuang United Technology Co., Ltd.

Quality Department

Sep 6, 2023